

## Jenna Leifso

---

**From:** Sahar, Zalma (MCSCS) <Zalma.Sahar@ontario.ca> on behalf of Pegg, Jon (MCSCS) <Jon.Pegg@ontario.ca>  
**Sent:** May-06-19 12:12 PM  
**To:** OFM Subscribers (MCSCS)  
**Subject:** Fire Marshal Notification to Fire Chiefs - Health Canada Notification - Illegal Marketing of Infrared Saunas to Fire Departments with False Cancer Claims  
**Attachments:** EN\_19-101504-117\_HCCorrespondence.pdf  
**Follow Up Flag:** Follow up  
**Flag Status:** Flagged

Dear Mayor and Members of Council and Fire Chiefs:

**Reference: Request from Health Canada to advise on matters related to infrared saunas.**

I am writing to provide for your consideration a copy of correspondence I received from Health Canada - Marketed Health Products Directorate. If you have any questions, or require additional information, you may wish to contact Health Canada directly.

Yours truly,

Jon

Attachment: Health Canada Letter – File 19-101504-117

Jon Pegg  
Fire Marshal and Chief, Emergency Management  
Office of the Fire Marshal and Emergency Management  
Ministry of the Solicitor General  
Forensic Services and Coroner's Complex  
25 Morton Shulman Avenue  
Toronto, ON, M3M 0B1  
Telephone: 647-329-1239



Health  
Canada

Health Products  
and Food Branch

Santé  
Canada

Direction générale des produits  
de santé et des aliments

Marketed Health Products Directorate  
A.L. 1912C  
OTTAWA, Ontario  
K1A 0K9

File: 19-101504-117

May 2, 2019

Dear Fire Marshals,

**Re: Illegal Marketing of Infrared Saunas to Fire Departments with False Cancer Prevention Claims**

Health Canada would like to alert you to the illegal sale and marketing of infrared saunas to fire departments for unproven and false medical claims such as prevention of cancer amongst fire fighters and detoxification of fire fighters. Health Canada has discovered that at least two Canadian fire departments have acquired them for the aforementioned purposes. We are also concerned that the issue may be more widespread.

Health Canada assesses devices for their safety, quality and effectiveness before issuing a licence. Devices that are not licensed have not been assessed. Health Canada requests that you inform the fire departments in your province that Health Canada has **not** authorized for sale any infrared saunas as medical devices. Health Canada also invites fire departments to report any marketing of infrared saunas for medical purposes that they may encounter at [drug-device-marketing@canada.ca](mailto:drug-device-marketing@canada.ca).

Please do not hesitate to contact us if you have any questions, or if you wish to confirm the authorized uses for any device that has been marketed to you for medical purposes.

Sincerely,

Alain G. Musende, PhD  
Manager, Section for Transparency and Advertising  
Regulatory Surveillance  
Marketed Health Products Directorate  
Tel: (613) 954-6780 Fax: (613) 948-7996

Canada